

**Notice of Allowability**

Application No.

10/801,583

Applicant(s)

SHENG, THOMAS

Examiner

Cheukfan Lee

Art Unit

2625

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--**

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to an application for patent filed March 17, 2004.
2. ☒ The allowed claim(s) is/are 1-20.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☒ All    b) ☐ Some\*    c) ☐ None    of the:
    1. ☒ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.  
**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
5. ☐ CORRECTED DRAWINGS ( as "replacement sheets") must be submitted.
  - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review ( PTO-948) attached
    - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
  - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.

Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

**Attachment(s)**

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☒ Information Disclosure Statements (PTO/SB/08),  
Paper No./Mail Date 3/19/2007
4. ☐ Examiner's Comment Regarding Requirement for Deposit  
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),  
Paper No./Mail Date \_\_\_\_\_.
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other \_\_\_\_\_.

  
Cheukfan Lee

Art Unit: 2625

1. Claims 1-20 are allowed. Claims 1 and 20 are independent.

2. The following is an examiner's statement of reasons for allowance:

Claim 1 is allowable over the prior art of record. Claim 1 requires an automatic document feeder for feeding a document across a scanning region on a sheet passageway for a scanning module to scan, wherein a section of a guide plate of the sheet passageway located within the scanning region is formed with a concave portion, and the document moving across the scanning region is located between the concave portion and the scanning module, such that the intensity of a first brightness of the concave portion sensed by the scanning module is different from the intensity of a second brightness of the document sensed by the scanning module, to facilitate detection of a document edge, in combination with other limitation of claim 1, in a sheet-fed scanning device capable of detecting a document edge.

Claims 2-11 depend on claim 1.

Claim 12 claims the limitations of claim 1 and additional limitations. Claim 12 is allowable for the reason given for claim 1.

Claims 13-20 depend on claim 12.

Note: Rubscha (US 6,166,394), cited below, discloses a scanning device capable of detecting a document edge, comprising an automatic document feeder having a document guide plate in form of a convex portion, wherein the intensity of

Art Unit: 2625

brightness of the a surface of the convex portion sensed by the scanning module (90) is different from the intensity of the brightness of the document sensed by the module (90), to facilitate the detection of the document edge. Rubscha does not disclose the claimed concave portion.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

3. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Rubscha (US 6,166,394)

Spencer (US 6839,155)

Ferrell (US 6,741,741), Figs. 1 and 2

4. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Cheukfan Lee whose telephone number is (571) 272-7407. The examiner can normally be reached on 9:30 a.m. to 6:00 p.m., Mon-Fri.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Edward L. Coles can be reached on (571) 272-7402. The fax phone

Art Unit: 2625

number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Cheukfan Lee  
September 30, 2007